

## Restriction on use / Precautions for use

- ⤿ The product must be used within the recommended weight limits and in accordance with the assessment made by an approved ortho-prosthetist depending on the patient concerned
- ⤿ Product maintenance is subject to an assessment made by an approved orthoprosthetist depending on the conditions of use
- ⤿ All assembly and adjustment work must be carried out by qualified personnel

## Cleaning / Storage

- ⤿ Stock the product in a dry place, in its original packing (on which the batch number is marked)
- ⤿ The frequency of inspections to check the equipment is subject to the assessment made by the orthoprosthetist depending on the patient concerned

## Guarantee

The guarantee is valid for one year as from the date of purchase of the product, if the product has been used in accordance with the manufacturer's instructions.

*N.B. In the event of a complaint, please enclose a return form correctly filled in, incomplete forms cannot be dealt with.*

## EC markings

Groupe G2M hereby declares, in its quality as the manufacturer, that all its parts comply with directive 93/42/EC.



L35

L37

## Désignation

- ⤿ L35/L36 : Dispositif de perlstein seul
- ⤿ L37/L38/L39 : Dispositif de perlstein avec tube

## Destination

Composant entrant dans la fabrication d'un pied d'orthèse du membre inférieur avec Perlstein.

## Pièces détachées

	L35	L36
Montants	L3501	L3601
Noix seule (rondelle taraudée M6)	L3502	L3602
Vis	V0098	V0099

	L37	L38	L39
Dispositif de Perlstein	L35	L35	L36
Diamètre du tube	8 mm	6 mm	6 mm
Épaisseur	11 mm	9 mm	9 mm
Longueur	100 mm	90 mm	90 mm

## Restriction d'utilisation / Précautions d'emploi

- Les appareils de Perlstein doivent être utilisés suivant l'appréciation de l'orthoprothésiste agréé en fonction du patient à appareiller
- La maintenance du produit reste à l'appréciation de l'orthoprothésiste selon le cadre de l'utilisation
- L'assemblage et l'ajustage doivent être effectués par un personnel qualifié

## Entretien / Stockage

- Stocker le produit à l'abri de l'humidité et dans son emballage d'origine (sur lequel est inscrit son numéro de lot)
- La fréquence des visites de contrôle de l'appareil reste à l'appréciation de l'orthoprothésiste en fonction du patient à appareiller

## Garantie

La garantie est valable 1 an pour une utilisation conforme aux instructions du fabricant à partir de la date d'achat du produit.

NB : Pour toutes réclamations veuillez joindre un formulaire de retour dûment complété, tout formulaire incomplet ne pourra être traité.

## Marquage CE

Groupe G2M déclare, en sa qualité de distributeur, que toutes ses pièces sont conformes à la directive 93/42/CE.

## L35/L36/L37/L38/L39



L35

L37

## Designation

- L35/L36 : Stainless steel Perlstein's apparatus only
- L37/L38/L39 : Stainless steel Perlstein's apparatus with tube

## Purpose

Component used for the manufacture of a foot orthosis for lower limbs, with Perlstein.

## Spares parts

	L35	L36
Plate only	L3501	L3601
Nut only (tapped washer M6)	L3502	L3602
Screw	V0098	V0099

	L37	L38	L39
Perlstein's apparatus	L35	L35	L36
Tube diameter	8 mm	6 mm	6 mm
Thick	11 mm	9 mm	9 mm
Long	100 mm	90 mm	90 mm